

2023, (including Minor Correction MC24002 applied May 17, 2024).

(v) WGQ Quadrant Electronic Delivery Mechanism Related Standards, Version 4.0, September 29, 2023, (including Revision to Standard No. 4.3.23, November 25, 2024).

(vi) WGQ Capacity Release Related Standards, Version 4.0, September 29, 2023, (including Addition of WGQ Standard No. 5.3.74, November 25, 2024).; and

(vii) WGQ Cybersecurity Related Standards, Version 4.0, September 29, 2023.

\* \* \* \* \*

[FR Doc. 2026-10600 Filed 5-27-26; 8:45 am]

BILLING CODE 6717-01-P

## DEPARTMENT OF HEALTH AND HUMAN SERVICES

### Food and Drug Administration

#### 21 CFR Part 876

[Docket No. FDA-2026-N-5200]

#### Medical Devices; Gastroenterology-Urology Devices; Classification of the Endoscopic Suturing Device for Altering Gastric Anatomy for Weight Loss

**AGENCY:** Food and Drug Administration, HHS.

**ACTION:** Final amendment; final order.

**SUMMARY:** The Food and Drug Administration (FDA) is classifying the endoscopic suturing device for altering gastric anatomy for weight loss into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for classification of the endoscopic suturing device for altering gastric anatomy for weight loss. We are taking this action because we have determined that classifying the device into class II will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices, in part by reducing regulatory burdens.

**DATES:** This order is effective May 28, 2026. The classification was applicable on July 12, 2022.

**FOR FURTHER INFORMATION CONTACT:** April Marrone, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 5449, Silver Spring, MD 20993-0002, 240-402-6510, [April.Marrone@fda.hhs.gov](mailto:April.Marrone@fda.hhs.gov).

**SUPPLEMENTARY INFORMATION:**

### I. Background

Upon request, FDA (the Agency or we) has classified the endoscopic suturing device for altering gastric anatomy for weight loss into class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness of the device. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified into, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act (see also part 860, subpart D (21 CFR part 860, subpart D)). Section 207 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105-115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112-144) modified the De Novo classification process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a premarket notification (510(k)) for a device that has not previously been classified. After

receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

We believe this De Novo classification will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens. When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining "substantial equivalence"). Instead, sponsors can use the less burdensome 510(k) process, when necessary, to market their device.

### II. De Novo Classification

On September 30, 2021, FDA received Apollo Endosurgery, Inc.'s request for De Novo classification of the APOLLO ESG System, APOLLO ESG SX System, APOLLO REVISE System, and APOLLO REVISE SX System devices. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable assurance of safety and effectiveness of the device, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see section 513(a)(1)(B) of the FD&C Act). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special

controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on July 12, 2022, FDA issued an order to the requester classifying the device into class II. In

this final order, FDA is codifying the classification of the device by adding 21 CFR 876.5983.<sup>1</sup> We have named the generic type of device “endoscopic suturing device for altering gastric anatomy for weight loss,” and it is identified as a device that uses suturing

to approximate gastric tissue to restrict the volume of the stomach for the intended purpose of weight loss.

FDA has identified the risks to health associated with this type of device and the measures required to mitigate these risks in table 1.

**TABLE 1—RISKS TO HEALTH AND MITIGATION MEASURES FOR ENDOSCOPIC SUTURING DEVICES FOR ALTERING GASTRIC ANATOMY FOR WEIGHT LOSS**

Identified risks to health	Mitigation measures
Device- and/or procedure-related adverse events, including: <ul style="list-style-type: none"> <li>• Death</li> <li>• Gastrointestinal bleeding</li> <li>• Obstruction</li> <li>• Perforation</li> <li>• Injury to organs adjacent to the stomach</li> <li>• Perigastric leak</li> <li>• Nausea</li> <li>• Infection</li> <li>• Pain</li> <li>• Pneumoperitoneum</li> <li>• Pneumothorax</li> <li>• Pulmonary embolism</li> </ul> Weight gain ..... Adverse tissue reaction .....	Clinical performance testing; Non-clinical performance testing; Labeling; Training; Sterilization validation; and Shelf life testing.  Clinical performance testing Labeling. Biocompatibility evaluation.

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness of the device. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this final order.

Under the FD&C Act, submission of a premarket notification under section 510(k) is required to reasonably assure the safety and effectiveness of class II devices unless FDA determines that the device type should be exempt under section 510(m) of the FD&C Act. At this time FDA has not made this determination for endoscopic suturing devices for altering gastric anatomy for weight loss. This device is therefore subject to premarket notification requirements under section 510(k) of the FD&C Act.

**III. Analysis of Environmental Impact**

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not normally have a significant effect on the human environment. Therefore, neither an environmental assessment nor an

environmental impact statement is required.

**IV. Paperwork Reduction Act of 1995**

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations and guidance. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501–3521). The collections of information in part 860, subpart D, regarding De Novo classification have been approved under OMB control number 0910–0844; the collections of information in 21 CFR part 814, subparts A through E, regarding premarket approval have been approved under OMB control number 0910–0231; the collections of information in part 807, subpart E, regarding premarket notification submissions have been approved under OMB control number 0910–0120; the collections of information in 21 CFR part 820 regarding quality management system regulation have been approved under OMB control number 0910–0073; and the collections of information in 21 CFR part 801 regarding labeling have been approved under OMB control number 0910–0485.

**List of Subjects in 21 CFR Part 876**

Medical devices. Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR Part 876 is amended as follows:

**PART 876—GASTROENTEROLOGY-UROLOGY DEVICES**

■ 1. The authority citation for part 876 continues to read as follows:

**Authority:** 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

■ 2. Add § 876.5983 to subpart F to read as follows:

**§ 876.5983 Endoscopic suturing device for altering gastric anatomy for weight loss.**

(a) *Identification.* An endoscopic suturing device for altering gastric anatomy for weight loss uses suturing to approximate gastric tissue to restrict the volume of the stomach for the intended purpose of weight loss.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Clinical performance testing must demonstrate the device performs as intended under anticipated conditions of use and evaluate the following:

- (i) Weight change; and
- (ii) All adverse events.

<sup>1</sup> FDA notes that the “ACTION” caption for this final order is styled as “Final amendment; final order,” rather than “Final order.” Beginning in December 2019, this editorial change was made to

indicate that the document “amends” the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register’s (OFR) interpretations of the Federal Register Act (44

U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

(2) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use. The following performance characteristics must be tested:

(i) Performance bench testing in a simulated use model must verify functional aspects of the device design and support device durability during clinical use;

(ii) Dimensional specifications must be verified; and

(iii) Tensile strength testing must be performed for all articulating components.

(3) Performance data must support the shelf life of the device by demonstrating continued package integrity and device functionality over the labeled shelf life.

(4) Performance data must demonstrate the sterility of the patient-contacting components of the device.

(5) The patient-contacting components of the device must be demonstrated to be biocompatible.

(6) Training must be provided so that, upon completion of the training program, the user can use the device correctly to approximate tissue to alter the gastric anatomy for the purpose of weight loss with minimal impact to the safety of the patient.

(7) Labeling must include:

(i) A summary of clinical performance testing with the device, including a discussion of adverse events and clinical benefit reported as percent total body weight loss; and

(ii) A shelf life.

**Grace R. Graham,**

*Deputy Commissioner for Policy, Legislation, and International Affairs.*

[FR Doc. 2026-10621 Filed 5-27-26; 8:45 am]

**BILLING CODE 4164-01-P**

## DEPARTMENT OF HOMELAND SECURITY

### Coast Guard

#### 33 CFR Part 100

[Docket No. USCG-2026-0598]

#### Special Local Regulations; Marine Events Within the Sector Columbia River Captain of the Port Zone

**AGENCY:** Coast Guard, DHS.

**ACTION:** Notification of enforcement of regulation.

**SUMMARY:** The Coast Guard will enforce special local regulations at various locations in the Sector Columbia River Captain of the Port Zone from May 30, 2026, to August 8, 2026. This action is

necessary to provide for the safety of life and property on these navigable waters during marine events. During the enforcement periods, the operator of any vessel in the regulated area must comply with the directions from the Patrol Commander or any official patrol vessel.

**DATES:** The regulations in 33 CFR 100.1302 will be enforced for the regulated areas identified in the **SUPPLEMENTARY INFORMATION** section below for the dates and times specified.

**FOR FURTHER INFORMATION CONTACT:** If you have questions about this notice of enforcement, call or email Lieutenant Commander Jesse Wallace, Waterways Management Division, Sector Columbia River, Coast Guard; telephone 503-572-3524, email [SCRWWM@USCG.MIL](mailto:SCRWWM@USCG.MIL).

**SUPPLEMENTARY INFORMATION:** The Coast Guard will enforce special local regulations in 33 CFR 100.1302 for the following events found in lines 1, 2, and 6 of Table 1 to § 100.1302. The event in line 1, the Rose Fest Dragon Boat Races, will be held from 7:00 a.m. to 4:00 p.m. on June 6, 2026 through June 7, 2026 in Portland, OR. Regulated area includes all waters of the Willamette River shore to shore, bordered on the north by the Hawthorne Bridge, and on the south by the Marquam Bridge.

The event in line 2, Spring Testing Hydroplane races, will be held on May 30, 2026, from 8:00 a.m. to 6:00 p.m. in Kennewick, WA. Regulated area includes all navigable waters within the Columbia River in the vicinity of Columbia Park, commencing at the Interstate 395 Bridge and continuing upriver approximately 2.0 miles and terminating at the northern end of Wade Island.

The event in line 6, Swim the Snake, will be held on August 8, 2026, from 11:00 a.m. to 1:00 p.m. in Perry, WA. Regulated area includes all navigable waters, bank-to-bank of the Snake River, 500 yards upstream and 500 yards downstream from the Washington State Highway 261 Bridge at the approximate position of 46°35'23" N; 118°13'10" W. All coordinates reference Datum NAD 1983.

During the enforcement periods, as reflected in § 100.1302, if you are the operator of a vessel in the regulated area you must comply with the lawful directions from the Patrol Commander or any official patrol vessel. Vessels may not transit the regulated areas without approval from the Patrol Commander. Vessels permitted to transit must operate at a no wake speed, in a manner which will not endanger participants or other crafts in the event. Spectators or other vessels shall not anchor, block,

loiter, or impede the transit of event participants or official patrol vessels in the regulated areas during the effective dates and times, or dates and times as modified through Local Notice to Mariners, unless authorized by an official patrol vessel. In addition to this notice of enforcement in the **Federal Register**, the Coast Guard will provide notification of these enforcement periods via the Local Notice to Mariners and marine information broadcasts.

Dated: May 22, 2026.

**Anthony R. Migliorini,**

*CAPT, U.S. Coast Guard, Captain of the Port Sector Columbia River.*

[FR Doc. 2026-10597 Filed 5-27-26; 8:45 am]

**BILLING CODE 9110-04-P**

## DEPARTMENT OF HOMELAND SECURITY

### Coast Guard

#### 33 CFR Part 100

[Docket Number USCG-2026-0096]

RIN 1625-AA08

#### Special Local Regulation; Sail Grand Prix, Upper Bay, New York City, NY

**AGENCY:** Coast Guard, Department of Homeland Security.

**ACTION:** Temporary final rule.

**SUMMARY:** The Coast Guard is establishing a temporary special local regulation (SLR) for certain waters of the Upper Bay of New York Harbor. The SLR is needed to provide for the safety of life to mitigate the dangers associated with a high-speed sailboat race. This proposed rulemaking temporarily prohibits persons and vessels from entering, transiting through, blocking, or loitering within the "Event Area", unless authorized by the Captain of the Port New York or a designated representative.

**DATES:** This rule is effective from noon on May 30, 2026, through 6:30 p.m. on May 31, 2026. The rule will only be enforced, however, from noon until 6:30 p.m. on Saturday May 30, 2026, and from noon until 6:30 p.m. on Sunday, May 31, 2026.

**ADDRESSES:** To view available documents go to <https://www.regulations.gov> and search for USCG-2026-0096.

**FOR FURTHER INFORMATION CONTACT:** If you have questions about this rule, contact MSTC Sean Fitzgerald, Sector New York Waterways Management Division, U.S. Coast Guard; telephone 718-801-2932, or email [SectorNYWWM@uscg.mil](mailto:SectorNYWWM@uscg.mil).