

(iii) A detailed summary of the device technical parameters and the typical course of treatment;

(iv) Device- and procedure-related adverse events pertinent to use of the device; and

(v) A shelf life for any sterile components.

(9) Patient labeling must include:

(i) Post-operative care instructions to avoid infection and inflammation of the surgical site;

(ii) Instructions to avoid overstimulation related nerve and tissue damage;

(iii) Instructions to avoid mechanical injuries to nerve and tissue caused by the implanted component;

(iv) Instructions for reprocessing/cleaning of any reusable components;

(v) Clinical performance reported by relevant subgroups;

(vi) The risks and benefits associated with use of the device;

(vii) Information on the typical course of treatment;

(viii) Instructions to avoid pain and discomfort; and

(ix) Instructions to avoid interference with other medical devices.

Grace R. Graham,

Deputy Commissioner for Policy, Legislation, and International Affairs.

[FR Doc. 2026-18428 Filed 9-9-26; 8:45 am]

BILLING CODE 4164-01-P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 880

[Docket No. FDA-2026-N-9909]

Medical Devices; General Hospital and Personal Use Devices; Classification of the Whole Room Microbial Reduction Device

AGENCY: Food and Drug Administration, HHS.

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA) is classifying the whole room microbial reduction device into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for classification of the whole room microbial reduction device. We are taking this action because we have determined that classifying the device into class II will provide a reasonable assurance of the safety and effectiveness of the device. We believe this action

will also enhance patients' access to beneficial innovative devices, in part by reducing regulatory burdens.

DATES: This order is effective September 10, 2026. The classification was applicable on September 1, 2023.

FOR FURTHER INFORMATION CONTACT:

Christopher Dugard, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 4640, Silver Spring, MD 20993-0002, 240-402-6031, Christopher.Dugard@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Upon request, FDA (the Agency or we) has classified the whole room microbial reduction device into class II (special controls), which we have determined will provide a reasonable assurance of the safety and effectiveness of the device. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified into, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act (see also part 860, subpart D (21 CFR part 860, subpart D)). Section 207 of the Food and Drug

Administration Modernization Act of 1997 (Pub. L. 105-115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112-144) modified the De Novo classification process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a premarket notification (510(k)) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

We believe this De Novo classification will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens. When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining "substantial equivalence"). Instead, sponsors can use the less burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On February 1, 2023, FDA received Xenex Disinfection Services, Inc.'s request for De Novo classification of the LightStrike+ device. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable

assurance of the safety and effectiveness of the device, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see section 513(a)(1)(B) of the FD&C Act). After review of the information submitted in the request, we determined that the device can be classified into class II

with the establishment of special controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device. Therefore, on September 1, 2023, FDA issued an order to the requester classifying the device into class II. In this final order, FDA is codifying the classification of the device by adding 21

CFR 880.6510.¹ We have named the generic type of device “whole room microbial reduction device,” and it is identified as a medical device to be used to reduce microbial load on medical device surfaces following cleaning and disinfection. FDA has identified the risks to health associated with this type of device and the measures required to mitigate these risks in table 1.

TABLE 1—RISKS TO HEALTH AND MITIGATION MEASURES FOR WHOLE ROOM MICROBIAL REDUCTION DEVICES

Identified Risks to Health	Mitigation Measures
Exposure to microbicidal agent, leading to skin and eye damage	Non-clinical performance testing; biocompatibility evaluation; software verification, validation, and hazard analysis; and labeling.
Respiratory mucous membrane irritation and pulmonary edema due to chemical exposure.	Non-clinical performance testing; and biocompatibility evaluation.
Patient cross-contamination due to device failure leading to inadequate microbial reduction.	Non-clinical performance testing; labeling; and software verification, validation, and hazard analysis.
Electrical shock	Electrical safety testing; non-clinical performance testing; and labeling.
Interference with other devices	Electromagnetic compatibility testing; electrical safety testing; and labeling.

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of the safety and effectiveness of the device. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this final order.

Under the FD&C Act, submission of a premarket notification under section 510(k) is required to reasonably assure the safety and effectiveness of class II devices unless FDA determines that the device type should be exempt under section 510(m) of the FD&C Act. At this time FDA has not made this determination for whole room microbial reduction devices. This device is therefore subject to premarket notification requirements under section 510(k) of the FD&C Act.

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not normally have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations and guidance. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501–3521). The collections of information in part 860, subpart D, regarding De Novo classification have been approved under OMB control number 0910–0844; the collections of information in 21 CFR part 814, subparts A through E, regarding premarket approval have been approved under OMB control number 0910–0231; the collections of information in part 807, subpart E, regarding premarket notification submissions have been approved under OMB control number 0910–0120; the collections of information in 21 CFR part 820 regarding quality management system regulation have been approved under OMB control number 0910–0073; and the collections of information in 21 CFR part 801 regarding labeling have been approved under OMB control number 0910–0485.

List of Subjects in 21 CFR Part 880

Medical devices. Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner

of Food and Drugs, 21 CFR part 880 is amended as follows:

PART 880—GENERAL HOSPITAL AND PERSONAL USE DEVICES

- 1. The authority citation for part 880 continues to read as follows:
Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.
- 2. Add § 880.6510 to subpart G to read as follows:

§ 880.6510 Whole room microbial reduction device.

- (a) *Identification.* A whole room microbial reduction device is a medical device to be used to reduce microbial load on medical device surfaces following cleaning and disinfection.
- (b) *Classification.* Class II (special controls). The special controls for this device are:
 - (1) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use. The following performance characteristics must be tested:
 - (i) Performance testing must demonstrate microbial log reduction of the demonstrated most resistant microorganism on medical device surfaces commensurate with the intended level of microbial reduction;
 - (ii) Simulated use testing must evaluate device performance under simulated worst-case use conditions

¹ FDA notes that the “ACTION” caption for this final order is styled as “Final amendment; final order,” rather than “Final order.” Beginning in December 2019, this editorial change was made to

indicate that the document “amends” the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register’s (OFR) interpretations of the Federal Register Act (44

U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

(e.g., soiling, room objects and surfaces, distances);

(iii) In-use testing must evaluate device performance under real-world use conditions;

(iv) Performance testing must demonstrate the photobiological safety of any lamps or lamp systems;

(v) Performance testing must evaluate safety features intended to prevent exposure and ensure that device operation can only occur in an unoccupied environment; and

(vi) Performance testing must characterize the long-term material compatibility of the microbiocidal agent on clinically relevant surfaces and/or devices.

(2) Biocompatibility testing must demonstrate safe residual levels of chemicals on medical devices surfaces and/or gaseous byproducts in air.

(3) Software verification, validation, and hazard analysis must be performed for any software components.

(4) Performance data must demonstrate the electromagnetic compatibility and electrical safety of the device.

(5) Labeling must include:

(i) Warnings and instructions to ensure the device is operated in an unoccupied environment;

(ii) Setup and positioning instructions; and

(iii) Information regarding material compatibility.

Grace R. Graham,

Deputy Commissioner for Policy, Legislation, and International Affairs.

[FR Doc. 2026-18426 Filed 9-9-26; 8:45 am]

BILLING CODE 4164-01-P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 892

[Docket No. FDA-2026-N-9904]

Medical Devices; Radiology Devices; Classification of the Vaginal Hydrogel Packing System

AGENCY: Food and Drug Administration, HHS.

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA) is classifying the vaginal hydrogel packing system into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for classification of the vaginal hydrogel

packing system. We are taking this action because we have determined that classifying the device into class II will provide a reasonable assurance of the safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices, in part by reducing regulatory burdens.

DATES: This order is effective September 10, 2026. The classification was applicable on August 22, 2023.

FOR FURTHER INFORMATION CONTACT: Justina Tam, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 3661, Silver Spring, MD 20993-0002, 240-402-4974, Justina.Tam@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Upon request, FDA (the Agency or we) has classified the vaginal hydrogel packing system into class II (special controls), which we have determined will provide a reasonable assurance of the safety and effectiveness of the device. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified into, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act (see also part 860, subpart D (21 CFR part 860, subpart D)). Section 207 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105-115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112-144) modified the De Novo classification process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a premarket notification (510(k)) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

We believe this De Novo classification will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens. When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining "substantial equivalence"). Instead, sponsors can use the less burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On August 26, 2022, FDA received BrachyFoam, Inc. (d/b/a Advaray)'s request for De Novo classification of the BrachyGel Vaginal Hydrogel Packing